

Certificate-Register-No.: Z – 07 – 146 – MP – E

Berlin Cert GmbHInstitute for Testing and Certification of Medical Devices
associated with the Technical University Berlin

certifies, that the following product of the company

SDD Seating & Design & Development Ltd.6 Edward St.
GB - Birmingham B1 2RX

fulfills the requirements of the named standards

Product: Moule Tec™ medical wheelchair cushion**Requirements:** Test method 11-10 08/2004 MDS-Hi
(examination of the microclimatic characteristics)
Devices against pressure ulcer
Class D (86,5 % rH): No relevant transport of VapourTest method 11-11 08/2004 MDS-Hi
(examination of the shear force and the pressure
relief capability)
Devices against pressure ulcer
Class C (4 %): Adequate pressure reliefThe conducted tests are documented in the test protocols
P-07-146-MP-PA088-10 and P-07-146-MP-PA088-11This certificate is not a confirmation, that the product complies with the German medical
device law or the directive 93/42/EEC

Issued on: 2007-11-13

